



## DECLARATION OF CONFORMITY

Winnocare Nordic ApS. Taarnborgvej 12C, 4220 Korsoer, Denmark. SRN: DK-MF-000002147., as the manufacturer of following product/equipment declare compliance with the applicable provisions of Regulation (EU) 2017/745 of 05 April 2017 and the following standards: EN ISO 10535:2006, EN ISO 14971:2019, EN ISO 3759, EN ISO 6330 and EN ISO 5077, EN ISO 13485:2016.

Brand:



Tradename:

Ergosling

Model/Type:

- Universal Basic
- Universal Basic with head support
- Universal Basic, mesh
- Universal Basic with head support, mesh

SWL:

275 kg

From serial number:

10001 - 10002

Product classification:

- Product group: Body-support units for hoists
- Usage: Reusable medical device
- Sterility: Nonsterile

Classification:

MDR Class 1 in accordance with Rule 1 / 13 of Annex VIII of Regulation (EU) 2017/745

UDI number:

574000810071, 574000810072, 574000810073, 574000810074, 574000810075, 574000810076, 574000810077, 574000810078, 574000810079, 574000810080, 574000810081, 574000810082, 574000810083, 574000810084, 574000810085, 574000810086, 574000810087, 574000810088, 574000810089, 574000810090, 574000810091, 574000810092, 574000810093, 574000810094.

Korsoer, 18 May 2021

Place and date

Richard Christiansen, Technical director,  
Winnocare Nordic ApS

Winnocare Nordic ApS, Taarnborgvej 12C, 4220 Korsoer, Denmark

